

Circular No. (6) 2020

Date: 17 MAR 2020

To All Medical Device Importers

Subject: Air Doctor product

The National Health Regulatory Authority would like to attract your attention that any medical claims regarding the "Air Doctor" product stating that it can be used for the prevention of airborne viruses and bacteria are not accurate and not scientifically proven. NHRA will take the necessary legal regulatory measures regarding the product, distributors, and advertisers.

Therefore, please obtain information from credible and official sources only.

Your cooperation is highly appreciated in improving health services in the Kingdom.

تعميم رقم (6) لسنة 2020

التاريخ: 17 مارس 2020م

إلى جميع المستوردين للأجهزة الطبية

الموضوع: منتج Air Doctor

تود الهيئة الوطنية لتنظيم المهن والخدمات الصحية لفت انتباهكم بأن الادعاءات الطبية حول منتج Air Doctor تنص على أنه يستخدم للوقاية من الفيروسات والبكتيريا المنقولة عبر الهواء ليست دقيقة وغير مثبتة علمياً، والهيئة ستتخذ كافة الإجراءات النظامية اللازمة حيال المنتج والشركة المسوقة والمعلنين.

وعليه، يجب استسقاء المعلومات الصحيحة من المصادر الرسمية بمملكة البحرين.

هذا ما لزم توضيحه، ولكم جزيل الشكر والتقدير على تعاونكم الدائم في الارتقاء بالخدمات الصحية بالمملكة.

الدكتورة مريم عذبي الجلاهمة
الرئيس التنفيذي