

Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	Rad-G devices
Device Model	27977
Serial number	600074128
Manufacturer	Masimo Corporation
Country of Origin	Mexico
Reference	Link
Reason of Alert	NHRA initiates this FSN due to identifying certain devices powering off and on without pressing the power button, an issue that can result in an unintentional change in the power state.
Action should be taken	Please stop using the above mentioned medical device and contact the authorized representative Wael Pharmacy Co. W.L.L at rithyl@waelpharmacy.com to take the necessary action for recall.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh