


## Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	Sastech Hygiene
Lot/Batch No.	All
Manufacturer	Sastech Detergent Manufacturing Co. W.L.L
Country of Origin	Bahrain
NHRA Registration	Not Registered
Device picture	
Reason of Recall	<p>NHRA initiates this FSN due to the risks associated with the use of the above-mentioned hand sanitizing gel as it contains of <b>Benzalkonium Chloride</b> which cause the following serious side effects:</p> <ul style="list-style-type: none"> <li>• Swelling.</li> <li>• Redness.</li> <li>• Very bad skin irritation.</li> </ul>
Action should be taken	<p>Please stop using the above-mentioned product and dispose of it. In case of any adverse events, please report to NHRA at <a href="mailto:medical_devices@nhra.bh">medical_devices@nhra.bh</a></p>

For more information please contact [Medical\\_Devices@nhra.bh](mailto:Medical_Devices@nhra.bh)