

Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	Examination Gloves
Lot No.	All
Manufacturer	Lotus
Country of Origin	P.C.R
Reference	NHRA
Device picture	
Reason of Recall	NHRA initiates this FSN due to the material used in the above-mentioned gloves is being used before which may affect the sterility of the gloves.
Action taken	Please stop using the defected gloves and contact the manufacturer Season international trading & industries at 1770 4647 to take the necessary action for return.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh