


## Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	Shield Hand Sanitizing Gel
Lot No.	All
Manufacturer	Bonn Medical Industries
Country of Origin	Saudi
Reference	<a href="https://twitter.com/Saudi_FDA/status/1238499107007799299">https://twitter.com/Saudi_FDA/status/1238499107007799299</a>
Device picture	
Reason of Recall	NHRA initiates this FSN due to high percentage of methyl alcohol (methanol) in the above-mentioned product, and exposure to high levels of this substance may cause health risks to the consumer.
Action taken	Please stop using the above-mentioned defected sanitizing gel and dispose of it.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact [Medical\\_Devices@nhra.bh](mailto:Medical_Devices@nhra.bh)