


## Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	Sailor Viking hand sanitizing gel
Lot No.	All
Manufacturer	Sailor Viking Cleaning and Cosmetic manufacturing Co
Country of Origin	Turkey
Reference	NHRA
Device picture	
Reason of Recall	NHRA initiates this FSN due to the use of <b>Triethanolamine</b> and <b>Triclosan</b> in the ingredients of the above-mentioned hand sanitizer which are banned by FDA due to the serious side effects when absorbed by the skin such as swelling, itching, redness, allergic reactions and skin ulcerations.
Action should be taken	Please stop using the above-mentioned hand sanitizing gel and contact the manufacturer to take the necessary action for return.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact [Medical\\_Devices@nhra.bh](mailto:Medical_Devices@nhra.bh)

## تحذير سلامة للأجهزة و المستلزمات الطبية

تود الهيئة الوطنية لتنظيم المهن و الخدمات الصحية تنبيه جميع المرافق الصحية الحكومية والخاصة و الممثلين المعتمدين والموزعين بخصوص سحب الجهاز الطبي أدناه:

تفاصيل الجهاز	
اسم الجهاز	Sailor Viking hand sanitizing gel
رقم التشغيل	سايلور فايكينج جل معقم لليدين جميع التشغيلات
المصنع	Sailor Viking Cleaning and Cosmetic manufacturing Co
بلد النشأ	تركيا
صورة الجهاز	
سبب السحب	تم اصدار تحذير سلامة للمنتج المذكور اعلاه بسبب احتوائه على مادتي تريكلوسان (Triclosan) وثلاثي ايثانولامين (Triethanolamine) المحظور استخدامهم عالميا من هيئة الغذاء و الدواء الأمريكيه FDA حيث يتسببون في حدوث اثار جانبية خطيره على الجلد منها التورم ، الحكة ، الاحمرار ، الحساسية والتقرحات الجلدية.
الاجراء الواجب اتخاذه	يجب وقف استخدام المنتج المذكور اعلاه و التواصل مع المستورد لاتخاذ الاجراءات اللازمة لسحب المنتج.

تعاونك يساهم في تحسين الخدمات الصحية في مملكة البحرين.

For more information please contact [Medical\\_Devices@nhra.bh](mailto:Medical_Devices@nhra.bh)